

Titan Works, Broombank Road
 Chesterfield Trading Estate
 Chesterfield S41 9QJ UK
 Tel: 01246 260062
 Fax: 01246 260844
 Tel :
 Fax :
 www.weightron.com



Members of the
 UK Weighing Federation
 Membership No. 1070

Tel :
 Fax :
 www.weightron.com

This Calibration Certificate is issued in accordance with the UKWF Code of Practise for the calibration of non automatic weighing machines. All Weights used are traceable to NPL.

Customer

P A MOODY RECYCLING LTD NEST ROAD GATESHEAD NEWCASTLE
NE10 0ES
LOCATION

Machine Details

Cust Ref	
Make	Bilanciai
Model	
Ser No	T160600498
Capacity	50000kg,(400kg)
e/d	20kg(20kg)
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1	TR11030	28-Jul-20
Cert Interval		12 months
Next Cert	October 2021	

Tolerance Range +/-	Kg
0 - 10,000	20
10,000 - 40,000	40
40,000 - 50,000	60

Comments : SV Job No: 107588

2	4	6	8
1	3	5	7
X			

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Ldev	Unloading	UDev
0	0	0		
2,000	2,000	0	2,000	0
10,000	10,000	0	10,000	0
20,000	20,000	0	20,000	0
25,440	25,440	0	25,440	0
35,440	35,440	0	35,440	0
45,440	45,440	0		

Repeatability

Load	Ind	Ind	Ind
45,440	45,440	45,440	45,440

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 6,800	0	2 6,800	0
3 6,800	0	4 6,800	0
5 6,800	0	6 6,800	0
7 6,800	0	8 6,800	0

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signed: Craia Hickson

Engineer: Craig Hickson

Date: 29-Sep-2020